



Product Service

CERTIFICATE

No. Z1A 14 08 12906 231

Holder of Certificate: **DÜPERTHAL SICHERHEITSTECHNIK GMBH & Co. KG**



DÜPERTHAL
SICHERHEITSTECHNIK

Frankenstraße 3
63791 Karlstein
GERMANY

Factory(ies): 29272

Certification Mark:



Product: **Safety cabinets**

Model(s): **HSC line, 39-HHBBtu-xyz(ss)
HSC UTS ergo line, 39-HHBBTu-xyz (ss)**

Parameters:

Key code Art.-no.:	HSC line	HSC UTS ergo line
HH-height	13,20	06,08
BB-width	12	11,14
T-depth	6,7	5,6
u-closing technique	0,2,3	7,9
x-colour description	0,4,5,6,7	0,4,5,6,7
yz-interior fitting	00-99	00-99
ss-special sign	R,B,US,UT,AS,AD	R,B,US,UT,AS,AD

Fire resistance of cabinets type 90.
The article number defines exactly the material and the size of the product.
A detailed description of the parameters can be found in the test report.

Tested according to:

DIN EN 14470-1:2004
PPP 52125A:2011
DIN EN 14727:2006
EK5/AK4 09-10:2009
ZEK 01.4-08
TRGS 510:2013

The product meets the safety and health requirements of the German Product Safety Act section 20 to 22 ProdSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. See also notes overleaf.

Test report no.: 713043145

Valid until: 2019-09-01

Date, 2014-09-04

(Edgar Harnisch)

Page 1 of 1

