



Product Service

# CERTIFICATE

No. Z1A 11 12 12906 210

**Holder of Certificate:** DÜPERTHAL SICHERHEITSTECHNIK  
GMBH & Co. KG



**DÜPERTHAL**

Frankenstraße 3  
63791 Karlstein  
GERMANY

**Factory(ies):**

62099

**Certification Mark:**



**Product:**

Safety cabinets

**Model(s):**

PREMIUM line, 29-HHBBTu-xyz(ss)  
PREMIUM line one, 29-HHBBTu-xyz (ss)

**Parameters:**

Key code Art.-No.

HH height, stands for number 20  
BB width, stands for number 06, 12  
T depth, stands for number 6, 7  
u door technique, stands for number 2, 4  
x colour, stands for number 0, 4, 5, 6, 7  
yz interior fitting, stands for number 00-99  
ss special sign, stands for L, R, B  
Fire resistance FWF 90

Fire resistance: FWF 90

The article number defines exactly material and size of the product. A detailed description of the parameters can be found in the test report.

**Tested according to:**

DIN EN 14470-1:2004  
DIN EN 14727:2006  
EK5/AK4 09-10:2009

The product meets the safety and health requirements of the German Product Safety Act section 20 to 22 ProdSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. See also notes overleaf.

**Test report no.:**

71398418

**Valid until:**

2016-12-27

**Date,** 2011-12-28

(Edgar Harnisch)



Page 1 of 1