



Tel: (+48) 22 673 04 08, 22 879 70 46  
Fax: (+48) 22 610 55 36, 22 610 81 01  
E-mail: [mpw@mpw.pl](mailto:mpw@mpw.pl), [www.mpw.pl](http://www.mpw.pl)



Management  
System

[www.tuv.com](http://www.tuv.com)  
ID 0000023838

## STATEMENT OF CONFORMITY

No 10.352RH.03/ENG

We, „MPW MED. INSTRUMENTS” SPÓŁDZIELNIA PRACY  
46 Boremłowska Street, 04-347 Warsaw,  
Poland

hereby declare under our sole responsibility that the following machine is in  
accordance with the regulations of the EU Directive 98/79/EC – *in vitro*  
diagnostic medical devices – due to its conception and design.

Machine:	<b>Refrigerated and heated laboratory centrifuge</b>
Type:	<b>MPW – 352RH</b>
Catalog No:	<b>10352RH</b>
Product classification on the basis of the Directive 98/79/EC	Non classified to list A or B and not for self-testing

Applied harmonized standards:

PN-EN ISO 13485:2012	PN-EN ISO 13485:2012 AC:2013-03
PN-EN 61010-1:2011	PN-EN 61010-2-101:2005
PN-EN 61010-2-020:2008	PN-EN ISO 14971:2012
PN-EN 61326-2-6:2006	PN-EN ISO 15223-1:2012E
PN-EN 13612:2006	PN-EN ISO 62366:2008
PN-EN ISO 18113-2:2011	PN-EN ISO 18113-3:2011

„MPW MED. INSTRUMENTS”  
SPÓŁDZIELNIA PRACY  
w Warszawie

Z-ca Prezesa  
ds. Technicznych

*inż. Daniel Maciejewski*

PREZES ZARZĄDU

*mgr Hanna Malczyńska*

Warszawa, 21.07.2014