



INTERNET: <http://www.mpw.pl>
E-mail: mpw@mpw.pl
Tel: (+48 22) 610-56-21, 673-04-08,
Fax: (+48 22) 610-55-36, 610-81-01

STATEMENT OF CONFORMITY

No 10.055.03 / ENG

We, "MPW Med. instruments"
04-347 Warsaw, 46 Boremlowska Str.
Poland

hereby declare that the following machine is in accordance with the regulations of the EU Directive 98/79/EC -in vitro diagnostic medical devices- due to its conception and design.

Machine:

Laboratory Centrifuge

Type:

MPW – 55

Catalog No:

10055

Product classification
on the basis of the Directive
98/79/EC Annex III p. 2÷5

Non classified to list A or B

Relevant EU Directives:

EU IVD Directive 98/79/EC
EU Low Voltage Directive 73/23/EEC
EU EMC Directive 89/336/EEC

Applied harmonized standards:

EN 61010-1
EN 61010-2-020
EN 50081
EN 50082-1
EN 55011

MPW Med-Instruments
SPÓŁDZIELNIA PRACY
(6) w Warszawie

Warsaw 24.11.2005

Z-ca Prezesa
d/s Technicznych
mgr inż. Paweł Różycki

PREZES ZARZĄDU SP. NT
mgr Tadeusz Gaszczyński
Board of Management