



Tel: (+48) 22 673 04 08, 22 879 70 46  
Fax: (+48) 22 610 55 36, 22 610 81 01  
E-mail: [mpw@mpw.pl](mailto:mpw@mpw.pl), [www.mpw.pl](http://www.mpw.pl)



- Certified Management System
- DIN EN ISO 9001
- DIN EN ISO 13485

## STATEMENT OF CONFORMITY

No 10.056.03/ENG

We, „MPW Med. instruments” Spółdzielnia Pracy  
46 Boremłowska Street, 04-347 Warsaw,  
Poland

hereby declare under our sole responsibility that the following machine is in accordance with the regulations of the EU Directive 98/79/EC – *in vitro* diagnostic medical devices – due to its conception and design.

Machine:	<b>Laboratory Centrifuge</b>
Type:	<b>MPW – 56</b>
Catalog No:	<b>10056</b>
Product classification on the basis of the Directive 98/79/EC	Non classified to list A or B and not for self-testing

Applied harmonized standards:

PN-EN ISO 13485:2005	PN-EN 61010-1:2011
PN-EN 61010-2-020:2008	PN-EN 61010-2-101:2005
PN-EN 61326-2-6:2006	PN-EN ISO 14971:2011
PN-EN 13612:2006	PN-EN 980:2010
PN-EN ISO 18113-3:2011	

**MPW** Med-instruments  
SPÓLDZIELNIA PRACY  
(1) w Warszawie

Warszawa, 15.03.2012